



CANNA DI PROVA A PISTONE
APPARECCHIO DI MISURA
(esempio)




CANNA DI PROVA A PISTONE



Materiale n° 1.7225 (42 CrMo 4)

o equivalente (R = 1000 MPa)


CANNA DI PROVA A PISTONE PORTACARTUCCIA



* diametro da aggiustare alla canna all'atto del montaggio

Cal.	5,6/16	6,3/10	6,3/12	6,3/14	6,3/16	6,8/11	6,8/18	9x17	10x18
L ₃	16,33	11,0	13,0	15,0	17,0	12,0	19,0	18,8	19,0
P ₁ =H ₂	0 5,76	0 6,35	0 6,35	0 6,35	0 6,35	0 6,90	0 6,90	0 9,60	10,05
R ₁	0 7,30	0 7,70	0 7,70	0 7,70	0 7,70	0 8,55	0 8,55	0 11,20	0 10,95
R	1,15	1,25	1,25	1,25	1,25	1,45	1,45	1,30	1,10

CANNA DI PROVA A PISTONE PISTONE



M_p = 80 ± 1,0 g

V _a cm ³	T mm
0,04	0,25 + 0,01
0,08	0,50 + 0,01
0,16	1,00 + 0,02
0,25	1,56 + 0,02
0,40	2,50 + 0,05
0,60	3,70 + 0,05
0,80	5,00 + 0,05